

Workshop on QA issues in Si Detectors
CERN, 17-18.5.2001

Strip quality tests used by ATLAS SCT

A.Chilingarov
Lancaster University, UK

Layout

1. Introduction.
2. Simple strip test.
3. Full strip test.
4. Conclusions.

1. Introduction

An essential part of the QA procedure used by the ATLAS SemiConductor Tracker (SCT) for Si strip detectors is the strip quality test. It exists in two versions. A Simple Test measures the values of the coupling capacitors and verifies their ability to withstand a DC voltage up to 100V. It also reveals the breaks and shorts in the external metal strips.

A more complicated, Full Strip Test, does the same but in addition measures the values of polysilicon bias resistors and reveals the breaks in the strip implants and/or in the bias resistors.

The Simple Test can be done in light with a bare unbiased detector put directly on the chuck of the Probe Station, and it requires a single probe needle. The Full Strip Test should be done in darkness with the detector under the bias. Therefore it requires a special fixture to provide the bias contact(s) travelling with the detector. A special lift-search method developed in Lancaster allows corrections for possible deviations of the detector mounted in the fixture from the ideal flatness. This guarantees good contact with minimal scratching of the probing pad.

2. Simple Strip Test


ATLAS SCT uses single sided *p-in-n* Si strip detectors. The backside of the detector represents the n^+ contact completely covered with aluminium. The p^+ strips are connected to the bias rail via polysilicon bias resistors of $\sim 1.5\text{M}\Omega$. The readout is made via coupling capacitors of $\sim 150\text{pF}$.

Under standard operation conditions both implant and readout strips are at nearly ground potential. However a current surge through the strip and corresponding voltage drop at the bias resistor can produce a considerable voltage difference over the coupling capacitors for a short time. The coupling capacitor insulation was specified to be able to withstand 100V during at least 1 second.



In the Simple Test a bare unbiased detector is put directly on the chuck of the Probe Station in full light. The DC current produced by 10V and 100V and the capacitance are measured between each readout strip and the backside contact.

If the DC current exceeds 50 nA at 10V the measurement at 100V is skipped and the strip is identified as having a “pinhole” defect. If the current above 50 nA appears only at 100V this is called “punch-through”.



The abnormally low capacitor value indicates the break of the metal strip; the high capacitor value for several neighbouring strips shows the short between them.





SCHEMATIC DIAGRAM
OF TEST SET-UP



Pre-series w31 20220900600017 Strip Scan / Lancaster 4.04.2000



Pre-series w31 20220900600017 Strip Scan / Lancaster 4.04.2000




3. Full Strip Test

In the Full Strip Test the detector is put in a fixture allowing its biasing when the detector moves with the chuck. Typically the fixture is a frame with flexible wires connected to the detector via small bonding interface board. The same measurements as in the Simple Test are made between the readout strip and the bias rail without light and with about half of the depletion voltage applied between the backside and the bias rail of the detector.



Under these conditions the implant strips and bias resistors are not connected all together as in the Simple Test. Therefore the serial resistor measured by the LCR bridge gives the value of the bias resistor for the strip. The break in the p^+ implant becomes detectable since it produces the same effect as the break in the readout strip.



In the Simple Test the capacitance measurement can be made at AC frequency chosen in a wide range: 100Hz-100kHz. For the Full Strip Test the low frequency of ~100Hz is preferable.

The detector in the fixture may be not as flat and parallel to the chuck surface as the bare detector put directly on the chuck. To minimise the chuck lift and to preserve good contact over all detector area the lift height for each strip is found individually based on the quality of the capacitance measurement.




SCHEMATIC DIAGRAM OF TEST SET-UP










w12/04: C_s for 3 frequencies/Lancaster/Jan-01



w12/04: C_s ratios for 3 frequencies/Lancaster/Jan-01







Strip Analysis Summary at Ramped Volts


University of Cambridge

Device: 20220900201323, Hamamatsu bmSiDetectorOut


Date: 19 Feb 2001, 15:13:08

Comments: Full Strip Test


Bad strips: 8 out of 768
STRIP QUALITY: 98.96 %



| Analysis | |
|---------------------------------------|-------------------------------------|
| <input type="button" value="Update"/> | <input type="button" value="QUIT"/> |
| Cap | Mean: 143.3 Sigma: 0.45 |
| RBias | 1.63 0.03 |
| Thresholds | |
| Low | High |
| Current: 50.0 nA | |
| Capacitance: 125.00 300.00 pF | |
| Resistance: 0.50 2.00 Mohm | |
| Strip Defect Lists: | |
| Pinholes: 0 | ** None ** |
| Punchthroughs: 0 | ** None ** |
| Shorts: 0 | ** None ** |
| Opens: 1 | 705 |
| Rbias: 0 | ** None ** |
| Implants: 7 | 241,372,379,558,569,584,635 |



Strip 372



Strip 379

Strip Analysis Summary at Ramped volts

University of Cambridge


Device: STN39229-00149, Hamamatsu bmSiDetectorOut

Date: 09 Aug 2000, 15:54:06

Comments: Full Strip Test

Bad strips: 6 out of 768

STRIP QUALITY: 99.22 %



Analysis

Update

QUIT


Cap Mean Sigma
152.6 0.42
RBias: 1.37 0.02

Thresholds



| | Low | High | |
|--------------|--------|--------|------|
| Current: | 50.0 | | nA |
| Capacitance: | 100.00 | 300.00 | pF |
| Resistance: | 0.50 | 2.00 | Mohm |

Strip Defect Lists:

| | | |
|----------------|---|------------------------------|
| Pinholes: | 0 | xxx None xxx |
| Punchthroughs: | 0 | xxx None xxx |
| Opens: | 6 | 595, 596, 597, 598, 599, 600 |
| Shorts: | 0 | xxx None xxx |
| Rbias: | 0 | xxx None xxx |
| Implants: | 0 | xxx None xxx |



Full strip test for sxx 37160-04/ Lancaster 26.01.2000



4. Conclusions

1. The strip quality tests used by ATLAS SCT measure the values and check the electric strength of the coupling capacitors, measure the values of the polysilicon bias resistors and reveal the defects in the readout strips, implant strips and bias resistors.
2. The Simple Strip Test can be done in light with a bare unbiased detector put directly on the chuck of a Probe Station. It is however insensitive to the defects in the strip implants and the bias resistors.
3. The Full Strip Test is performed in darkness with the detector put in a fixture allowing its biasing. It is more informative but technically more difficult than the Simple Test.
4. A lift-search procedure developed at Lancaster University allows corrections for the distortions of the detector flatness induced by the test fixture.